



Weigh Calibration Test Report

Client **Scaletec**

Cert# 1493

Make	Avery	Capacity	250 x 0.5kg	ID	
Model	3205CLE	Tare	+125kg	Location	Factory
Serial#	B642928			Condition	Fair

Applied Mass (kg)	Pre-Adjustment (kg)	Post-Adjustment (kg)
0	0	0
10	10.0	10.0
20	20.5	20.0
30	30.5	30.0
40	41.0	40.0
50	51.0	50.0
100	102.0	100.0
150	153.0	150.0
200	203.0	200.0
250	250.0	250.0

Corner test (kg)	
60.0	60.0
60.0	
60.0	60.0

Comments
SAMPLE ONLY - Rebalance needle assembly.

Date Tested

Retest Date

Pass

Completed by: Tony Jarman

Signed: 

All masses used and test results shown are traceable to the national standard.
Masses test cert# W1863 issued by Ministry of consumer affairs.